CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: E2630054

CSI Calibration Number: 190328401
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 28-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: E2630054

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.7	20.2	-0.5	19.3	18.6	-0.7	±2
50.0	50.3	0.3	50.0	50.3	0.3	±2
87.7	88.1	0.4	87.7	88.0	0.3	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.9	25.0	0.1	26.4	26.5	0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
26.6	22.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: E3010006

CSI Calibration Number: 190322642
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: E3010006

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	20.1	-0.3	20.4	20.1	-0.3	±2
49.8	50.3	0.5	49.8	50.3	0.5	±2
90.2	90.5	0.3	90.0	90.3	0.3	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	24.2	0.2	24.0	24.2	0.2	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	23.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: Y2410067

CSI Calibration Number: 190322644
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: Y2410067

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	20.1	-0.3	20.4	20.1	-0.3	±2
49.8	50.3	0.5	49.8	50.3	0.5	±2
90.2	90.3	0.1	90.0	90.1	0.1	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	24.1	0.1	24.0	24.1	0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	23.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: X2120034

CSI Calibration Number: 190328403
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 28-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: X2120034

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.7	20.1	-0.6	19.3	18.8	-0.5	±2
50.0	50.4	0.4	50.5	50.6	0.1	±2
87.7	88.8	1.1	89.0	88.8	-0.2	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.9	25.1	0.2	26.4	26.4	0.0	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
26.6	22.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: X3410053

CSI Calibration Number: 190322644
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: X3410053

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	19.8	-0.6	20.7	20.2	-0.5	±2
49.8	49.9	0.1	50.1	50.2	0.1	±2
90.2	90.2	0.0	88.6	88.6	0.0	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	24.2	0.2	25.3	25.5	0.2	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	24.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: X3410060

CSI Calibration Number: 190322646
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: X3410060

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	19.9	-0.5	20.4	19.9	-0.5	±2
49.8	50.2	0.4	49.8	50.2	0.4	±2
90.2	90.6	0.4	90.0	90.4	0.4	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	24.0	0.0	24.0	24.0	0.0	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	23.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: B1240025

CSI Calibration Number: 190322644
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: B1240025

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	19.8	-0.6	20.4	19.8	-0.6	±2
49.8	50.1	0.3	49.8	50.1	0.3	±2
90.2	90.3	0.1	90.0	90.2	0.2	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	24.0	0.0	24.0	24.0	0.0	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	23.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: A2220055

CSI Calibration Number: 190328593
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 28-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: A2220055

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.5	19.9	-0.6	20.5	19.9	-0.6	±2
50.2	50.5	0.3	50.2	50.5	0.3	±2
89.0	89.1	0.1	89.0	89.0	0.0	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
26.0	26.0	0.0	26.0	26.0	0.0	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
25.3	20.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: D3520074

CSI Calibration Number: 190322648
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: D3520074

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	19.7	-0.7	20.4	19.7	-0.7	±2
49.8	49.9	0.1	49.8	49.9	0.1	±2
90.2	90.7	0.5	90.0	90.6	0.6	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	23.9	-0.1	24.0	23.9	-0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	23.

Brad Mbach

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: X3410056

CSI Calibration Number: 190328409
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 28-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: X3410056

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.7	20.7	0.0	19.3	18.8	-0.5	±2
50.0	51.1	1.1	50.5	50.8	0.3	±2
87.7	89.0	1.3	89.0	88.7	-0.3	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.9	25.1	0.1	26.4	26.3	-0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
26.6	22.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: D3520028

CSI Calibration Number: 190322644
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: D3520028

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	19.9	-0.5	20.4	19.9	-0.5	±2
49.8	50.1	0.3	49.8	50.0	0.2	±2
90.2	90.7	0.5	90.0	90.6	0.6	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	24.1	0.1	24.0	24.1	0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	23.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: D3520017

CSI Calibration Number: 190328598
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 28-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: D3520017

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.7	20.7	0.0	20.5	20.3	-0.2	±2
50.0	51.1	1.1	50.2	50.5	0.3	±2
87.7	88.8	1.1	89.0	88.7	-0.3	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.9	25.0	0.1	26.0	26.1	0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
25.3	20.

Brady Mikell

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: C4210024

CSI Calibration Number: 190328405
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 28-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: C4210024

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.7	20.0	-0.7	19.3	18.9	-0.4	±2
50.0	50.2	0.2	50.5	50.9	0.4	±2
87.7	88.3	0.6	89.0	88.8	-0.2	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.9	25.1	0.2	26.4	26.5	0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
26.6	22.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: X3410059

CSI Calibration Number: 190328591
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 28-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: X3410059

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.5	20.0	-0.5	20.5	20.0	-0.5	±2
50.2	50.5	0.3	50.2	50.5	0.3	±2
89.0	89.1	0.1	89.0	89.1	0.1	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
26.0	26.0	0.0	26.0	26.0	0.0	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
25.3	20.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: F4010005

CSI Calibration Number: 190328409
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 28-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: F4010005

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.7	20.5	-0.2	19.3	18.9	-0.4	±2
50.0	50.6	0.6	50.0	50.6	0.6	±2
87.7	88.2	0.5	87.7	88.1	0.4	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.9	24.9	0.0	26.4	26.4	0.0	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
26.6	22.

Brad Mikesell

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP45C Serial Number: E2630055

CSI Calibration Number: 190322644
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: E2630055

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	20.2	-0.2	20.7	20.4	-0.3	±2
49.8	50.5	0.7	50.1	50.3	0.2	±2
90.2	90.8	0.6	88.6	88.3	-0.3	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	24.1	0.1	25.3	25.4	0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	24.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP155 Serial Number: F3810057

CSI Calibration Number: 190321693
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: F3810057

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.3	20.7	0.4	20.5	20.9	0.4	±2
49.7	50.1	0.4	50.3	50.3	0.0	±2
89.1	89.9	0.8	89.2	89.7	0.5	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.5	24.6	0.1	25.0	25.1	0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.7	17.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP155 Serial Number: F3810056

CSI Calibration Number: 190322648
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: F3810056

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	20.8	0.4	20.7	21.0	0.3	±2
49.8	50.6	0.8	50.1	50.5	0.4	±2
90.2	91.7	1.5	88.6	89.2	0.6	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	23.9	-0.1	25.3	25.2	-0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	24.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP155 Serial Number: F3810047

CSI Calibration Number: 190321695
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

Instrument Calibration Condition

Received Disposition: In Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: F3810047

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.3	20.6	0.3	20.6	21.1	0.5	±2
49.7	50.2	0.5	50.3	50.9	0.6	±2
89.1	89.8	0.7	89.2	89.6	0.4	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.5	24.5	0.0	24.7	24.8	0.1	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.7	17.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

CAMPBELL SCIENTIFIC
WHEN MEASUREMENTS MATTER

Certificate of Calibration

Model: HMP155 Serial Number: H2940034

CSI Calibration Number: 190322642
Calibration Procedures: TST_HMPCAL R4 PRC33A R24

(Imelyak)
2017-2018

Instrument Calibration Condition

Received Disposition: Out of Tolerance
Returned Disposition: In Tolerance

Recommended Calibration Schedule

If the customer has not requested a calibration interval, a non-mandatory recommended interval is provided. Based on past experience and assumed normal usage, it is recommended that this instrument be calibrated in accordance with the interval stated below to insure sustained accuracy and reliable performance.

Calibration Date: 21-Mar-19 **Recommended Interval:** 1 Year of Service

Report of Calibration Standards Used

Make/Model	Serial Number	Cal. Due Date	Trace Number
HMT-337	F0840130	05-Jul-19	2000939-180705-HMT337-F0840130
CR3000	1870	28-Feb-22	170306585

CSI certifies the above instrument meets or exceeds published specifications and has been calibrated using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology, an accepted value of a natural physical constant or a ratio calibration technique. Policies and procedures at this facility comply with ISO-9001.

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F | www.campbellsci.com

Instrument Data Report

S/N: H2940034

Relative Humidity Measurements

As Received (%RH)			As Returned (%RH)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (%RH)**
20.4	21.1	0.7	20.7	20.8	0.1	±2
49.8	51.0	1.2	50.1	49.9	-0.2	±2
90.2	92.8	2.6	88.6	89.0	0.4	±2

Temperature Measurements

As Received (°C)			As Returned (°C)			
Reference	**Device**	**Difference**	**Reference**	**Device**	**Difference**	**Tolerance (°C)**
24.0	24.3	0.3	25.3	25.5	0.2	±0.2

Laboratory Environment

Temperature (°C)	Relative Humidity (%)
24.9	24.

Calibration Technician

Document shall not be reproduced except in full, without the written approval of Campbell Scientific, Inc.