



Xylem operates under
the brands of YSI,
SonTek, Aanderaa,
MJK, HYPACK and
SI Analytics.

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 1 of 8

Delivery Number	Packing Slip Number	Date Shipped
1814135	YSI-468698	13-FEB-17
Shipped Via: FEDEX-Air-FedEx 2nd Day		

CERTIFICATE OF COMPLIANCE

This is to certify that the materials, processes and finished products were controlled, tested, and accepted in accordance with our quality system and the applicable specifications.

Ron Geis
Vice President & General Manager

SOLD TO

NATIONAL PARK SERVICE
3100 NATIONAL PARK RD
JUNEAU, AK 99801, United States

SHIP TO

NATIONAL PARK SERVICE
3100 NATIONAL PARK RD
JUNEAU, AK 99801, United States

PICKUP DATE	PAYMENT TERMS	FREIGHT TERMS	SR NUMBER	INCOTERMS	TAX NAME	TAX NO.	Supplier No.
13-FEB-17	Credit Card	Collect	324411	2010 FCA - ORIGIN	Exempt		

ORDER LINE NO.	YSI ITEM NO./ HS NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	Lot# / Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
10.1	R6150 9801.10.0000	670869		CREDITC ARD		Repair, Model 6150 6150 ROX ODO PROBE SN# 08H101103 REPORTED SYMPTOM: PM TESTED AND CALIBRATED PROBE. UPDATE MOTOR ASSEMBLY/ALL UPDATES IN PLACE REPLACED QUAD SEAL. REPLACED O-RINGS. REPLACED MEMBRANE WITH PART#606155, LOT# 17A100600 YSI RECOMMENDS YEARLY		EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under the brands of YSI, SonTek, Aanderaa, MJK, HYPACK and SI Analytics.

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 2 of 8

Delivery Number	Packing Slip Number	Date Shipped
1814135	YSI-468698	13-FEB-17
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO./ HS NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	Lot# / Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						MEMBRANE REPLACEMENT. TESTED WIPER FUNCTION, NO DEFECT FOUND. PROBE READS: 100% IN AERATED BATH. PROBE READS: 1.2% IN ZERO TEST. FINAL TEST, MEETS YSI SPECS.					
12.1	R6920V2 9801.10.0000	670869		CREDITC ARD		Repair, Model 6920V2 MODEL# 6920V2 SN# 08H100354 REPORTED SYMPTOM: PM TESTED AND CALIBRATED BOARD, NO DEFECT FOUND. TESTED RTC BATTERY, NO DEFECT FOUND. UPDATE FIRMWARE TO LATEST VERSION (3.10)/FIRMWARE UP TO DATE AT 3.10. PERFORMED CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING. TESTED BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND. TESTED BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND. TESTED CONDUCTIVITY/TEMPERATURE PROBE (SN#13A100461)TEMPERATURE	United States	EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under
the brands of YSI,
SonTek, Aanderaa,
MJK, HYPACK and
SI Analytics.

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 3 of 8

Delivery Number	Packing Slip Number	Date Shipped
1814135	YSI-468698	13-FEB-17
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO./ HS NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	Lot# / Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						READS 20.8°C IN A 21.4°C BATH. OUT OF SPEC. **DO NOT REPLACE PER CUSTOMER REQUEST.** **PROBE WAS SCRAPPED AND PORT PLUG INSTALLED** TESTED PH PROBE(SN#16B) OUT OF SPEC. - LOW SLOPE. REPLACED PROBE WITH SN# 17A AND TESTED. READINGS BELOW. 7PH: - 1.6MV 10PH: - 177.6MV 4PH: 172.6MV SLOPE: 176.0 REPLACED O-RINGS AS NEEDED. FINAL TEST, MEETS YSI SPECS.					
14.1	R6150 9801.10.0000	670869		CREDITC ARD		Repair, Model 6150 6150 ROX ODO PROBE SN# 11E101714 REPORTED SYMPTOM: PM TESTED AND CALIBRATED PROBE. UPDATE MOTOR ASSEMBLY/ALL UPDATES IN PLACE REPLACED QUAD SEAL. REPLACED O-RINGS. REPLACED MEMBRANE WITH PART#606155, LOT# 17A100600 YSI RECOMMENDS YEARLY MEMBRANE REPLACEMENT. TESTED WIPER FUNCTION, NO		EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under
the brands of YSI,
SonTek, Aanderaa,
MJK, HYPACK and
SI Analytics.

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List


Page 4 of 8

Delivery Number	Packing Slip Number	Date Shipped
1814135	YSI-468698	13-FEB-17
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO./ HS NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	Lot# / Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						DEFECT FOUND. PROBE READS: 100% IN AERATED BATH. PROBE READS: 0.4% IN ZERO TEST. OUT OF SPEC. FINAL TEST, MEETS YSI SPECS.					
2.1	R6920V2 9801.10.0000	670869		CREDITC ARD		Repair, Model 6920V2 MODEL# 6920V2 SN# 08H100355 REPORTED SYMPTOM: ANNUAL MAINT. TESTED AND CALIBRATED BOARD, NO DEFECT FOUND. TESTED RTC BATTERY, NO DEFECT FOUND. UPDATE FIRMWARE TO LATEST VERSION (3.10)/FIRMWARE UP TO DATE AT 3.10. PERFORMED CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING. TESTED BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND. TESTED BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND. TESTED CONDUCTIVITY/TEMPERATURE PROBE (SN# 16A100578) TEMPERATURE READS 22.8°C IN A	United States	EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under
the brands of YSI,
SonTek, Aanderaa,
MJK, HYPACK and
SI Analytics.

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 5 of 8

Delivery Number	Packing Slip Number	Date Shipped
1814135	YSI-468698	13-FEB-17
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO./ HS NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	Lot# / Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						22.8°C BATH WITH A CELL CONSTANT OF 4.95046. PROBE READS ZERO SPCOND. IN AIR. TESTED PH PROBE(SN#15A) REPLACED PROBE WITH SN# 16M READINGS BELOW: 7PH: -7.6MV 10PH: -183.7MV 4PH: 169.5MV 176.1MV REPLACED O-RINGS AS NEEDED. FINAL TEST, MEETS YSI SPECS.					
4.1	R6150 9801.10.0000	670869		CREDITC ARD		Repair, Model 6150 6150 ROX ODO PROBE SN# 10H101536 REPORTED SYMPTOM: PREV. MAINT. TESTED AND CALIBRATED PROBE. UPDATE MOTOR ASSEMBLY/ALL UPDATES IN PLACE REPLACED QUAD SEAL. REPLACED O-RINGS. REPLACED MEMBRANE WITH PART#606155, LOT# 17A100600 YSI RECOMMENDS YEARLY MEMBRANE REPLACEMENT. TESTED WIPER FUNCTION, NO DEFECT FOUND. PROBE READS: 100% IN AERATED BATH.		EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under
the brands of YSI,
SonTek, Aanderaa,
MJK, HYPACK and
SI Analytics.

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 6 of 8

Delivery Number	Packing Slip Number	Date Shipped
1814135	YSI-468698	13-FEB-17
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO./ HS NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	Lot# / Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						PROBE READS: 1.4% IN ZERO TEST. FINAL TEST, MEETS YSI SPECS.					
6.1	R6136 9801.10.0000	670869		CREDITC ARD		Repair, Model 6136 6136 TURBIDITY PROBE SN# 11J100156 REPORTED SYMPTOM: PREV. MAINT. TESTED AND CALIBRATED PROBE. UPDATE MOTOR ASSEMBLY/ALL UPDATES IN PLACE REPLACED QUAD SEAL. REPLACED O-RINGS. TESTED WIPER FUNCTION, NO DEFECT FOUND. PROBE READS: 126 IN 126 STANDARD. FINAL TEST, MEETS YSI SPECS.		EA	1	1	0
8.1	R6920V2 9801.10.0000	670869		CREDITC ARD		Repair, Model 6920V2 MODEL# 6920V2 SN# 08H100353 REPORTED SYMPTOM: PM TESTED AND CALIBRATED BOARD, NO DEFECT FOUND. TESTED RTC BATTERY, NO DEFECT FOUND. UPDATE FIRMWARE TO LATEST VERSION (3.10)/FIRMWARE UP TO DATE AT 3.10. PERFORMED CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING.	United States	EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under
the brands of YSI,
SonTek, Aanderaa,
MJK, HYPACK and
SI Analytics.

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 7 of 8

Delivery Number	Packing Slip Number	Date Shipped
1814135	YSI-468698	13-FEB-17
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO./ HS NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	Lot# / Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						TESTED BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND.					
						TESTED BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND.					
						TESTED CONDUCTIVITY/TEMPERATURE PROBE (SN#16A100578)TEMPERATURE READS 22.3°C IN A 22.2°C BATH WITH A CELL CONSTANT OF 5.04657. PROBE READS ZERO SPCOND. IN AIR.					
						TESTED PH PROBE(SN#16A) OUT OF SPECS. REPLACED PROBE WITH SN# 17A AND TESTED. READINGS BELOW. 7PH: - 2.2MV 10PH: - 179.3MV 4PH: 171.7MV SLOPE: 176.1					
						REPLACED O-RINGS AS NEEDED. FINAL TEST, MEETS YSI SPECS.					

GROSS WEIGHT(LB) : 20.00 NET WEIGHT (LB) : 20.00 VOLUME(CC) : 0 TOTAL CARTONS: 0

ADDITIONAL INFO: 715018811893

ADDITIONAL NOTE: Shipping Note: Invoice Note:

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.



AAANDERAA®

mjk



SI Analytics



Xylem operates under
the brands of YSI,
SonTek, Aanderaa,
MJK, HYPACK and
SI Analytics.

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 8 of 8

Delivery Number	Packing Slip Number	Date Shipped
1814135	YSI-468698	13-FEB-17
Shipped Via: FEDEX-Air-FedEx 2nd Day		

Tracking Number:



-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

